

SUPERVISOR TEST

HUDSON FALLS  
VILLAGE POLICE DEPT.  
DATAMASTER dmt: 102408

Date: 02/22/2022

OPERATOR NAME:  
LANCE M BARBER  
Permit #: 42841  
Operator Agency/Dept: HUDSON FALLS VIL. PD  
Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:16
INTERNAL STANDARD	VERIFIED	00:17
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	00:17
BLANK TEST	0.00	00:18
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	00:18
BLANK TEST	0.00	00:19
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	00:19
BLANK TEST	0.00	00:20

