

SUPERVISOR TEST

HUDSON FALLS  
VILLAGE POLICE DEPT.  
DATAMASTER dmt: 102408

Date: 02/14/2022

OPERATOR NAME:  
DAVID M COSTELLO  
Permit #: 40891  
Operator Agency/Dept: HUDSON FALLS VIL. PD  
Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:12
INTERNAL STANDARD	VERIFIED	08:12
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	08:12
BLANK TEST	0.00	08:13
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	08:13
BLANK TEST	0.00	08:14
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	08:14
BLANK TEST	0.00	08:15

