

SUPERVISOR TEST

WHITEHALL VILLAGE  
POLICE DEPT.  
DATAMASTER dmt: 113806

Date: 02/12/2022

OPERATOR NAME:  
ALEXANDER T GILMORE  
Permit #: 47649  
Operator Agency/Dept: WHITEHALL VILL. P.D.  
Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:56
INTERNAL STANDARD	VERIFIED	21:56
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	21:56
BLANK TEST	0.00	21:57
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	21:58
BLANK TEST	0.00	21:58
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	21:59
BLANK TEST	0.00	22:00

