

SUPERVISOR TEST

WHITEHALL VILLAGE
POLICE DEPT.
DATAMASTER dmt: 113806

Date: 02/26/2022

OPERATOR NAME:
ALEXANDER T GILMORE
Permit #: 47649
Operator Agency/Dept: WHITEHALL VILL. P.D.
Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:48
INTERNAL STANDARD	VERIFIED	21:48
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	21:48
BLANK TEST	0.00	21:49
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	21:50
BLANK TEST	0.00	21:51
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	21:51
BLANK TEST	0.00	21:52

