

SUPERVISOR TEST

WHITEHALL VILLAGE  
POLICE DEPT.  
DATAMASTER dmt: 113806

Date: 02/19/2022

OPERATOR NAME:  
BRYAN J GRECO  
Permit #: 44077  
Operator Agency/Dept: WHITEHALL VILL. P.D.  
Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:14
INTERNAL STANDARD	VERIFIED	22:14
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	22:14
BLANK TEST	0.00	22:15
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	22:15
BLANK TEST	0.00	22:16
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	22:17
BLANK TEST	0.00	22:17

