

SUPERVISOR TEST

FORT EDWARD  
VILLAGE POLICE DEPT.  
DATAMASTER dmt: 115708

Date: 02/22/2022

OPERATOR NAME:  
JOSEPH P DEANGELO  
Permit #: 49218  
Operator Agency/Dept: FORT EDWARD VILL. PD  
Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:54
INTERNAL STANDARD	VERIFIED	15:54
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	15:54
BLANK TEST	0.00	15:55
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	15:56
BLANK TEST	0.00	15:57
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	15:57
BLANK TEST	0.00	15:58

JD  
#1217

