

SUPERVISOR TEST

 CAMBRIDGE VILLAGE
 POLICE DEPT
 DATAMASTER dmt: 126706

Date: 02/14/2022

OPERATOR NAME:
 ROBERT M DANKO
 Permit #: 39105
 Operator Agency/Dept: CAMBRIDGE VILL. PD
 Reference Standard #: 21390

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:12
INTERNAL STANDARD	VERIFIED	07:12
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	07:13
BLANK TEST	0.00	07:14
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	07:14
BLANK TEST	0.00	07:15
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	07:15
BLANK TEST	0.00	07:16

