

SUPERVISOR TEST

-----  
GRANVILLE VILLAGE  
POLICE DEPT  
DATAMASTER dmt: 113306  
-----

Date: 02/14/2022

OPERATOR NAME:  
THOMAS J ZOVISTOSKI  
Permit #: 36771  
Operator Agency/Dept: GRANVILLE VILLAGE PD  
Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:00
INTERNAL STANDARD	VERIFIED	04:00
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	04:00
BLANK TEST	0.00	04:01
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	04:01
BLANK TEST	0.00	04:02
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	04:02
BLANK TEST	0.00	04:03

