

SUPERVISOR TEST

WASHINGTON COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106406

Date: 02/14/2022

OPERATOR NAME:  
PAUL S HENRY  
Permit #: 49918  
Operator Agency/Dept: WASHINGTON CO. SO  
Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:24
INTERNAL STANDARD	VERIFIED	20:24
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	20:24
BLANK TEST	0.00	20:25
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	20:26
BLANK TEST	0.00	20:27
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	20:27
BLANK TEST	0.00	20:28

*Paul Henry #640*

