

SUPERVISOR TEST

WASHINGTON COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 106406

Date: 02/28/2022

OPERATOR NAME:
DALE W QUESNEL
Permit #: 37510
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:07
INTERNAL STANDARD	VERIFIED	11:07
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	11:07
BLANK TEST	0.00	11:08
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	11:09
BLANK TEST	0.00	11:09
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	11:10
BLANK TEST	0.00	11:11

