

SUPERVISOR TEST

WASHINGTON CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 100284

Date: 02/24/2022

OPERATOR NAME:
ERIC C WRIGHT
Permit #: 49047
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:25
INTERNAL STANDARD	VERIFIED	01:25
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	01:25
BLANK TEST	0.00	01:26
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	01:26
BLANK TEST	0.00	01:27
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	01:28
BLANK TEST	0.00	01:28

