

SUPERVISOR TEST

WASHINGTON CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 100284

Date: 02/28/2022

OPERATOR NAME:
ROBERT J SULLIVAN
Permit #: 34106
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:54
INTERNAL STANDARD	VERIFIED	06:55
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:55
BLANK TEST	0.00	06:56
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:56
BLANK TEST	0.00	06:57
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:57
BLANK TEST	0.00	06:58

