

SUPERVISOR TEST

HUDSON FALLS
VILLAGE POLICE DEPT.
DATAMASTER dmt: 102408

Date: 03/04/2022

OPERATOR NAME:
SEAN T CASEY
Permit #: 48101
Operator Agency/Dept: HUDSON FALLS VIL. PD
Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:47
INTERNAL STANDARD	VERIFIED	00:47
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	00:47
BLANK TEST	0.00	00:48
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	00:48
BLANK TEST	0.00	00:49
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	00:50
BLANK TEST	0.00	00:50

