

SUPERVISOR TEST

HUDSON FALLS
VILLAGE POLICE DEPT.
DATAMASTER dmt: 102408

Date: 03/16/2022

OPERATOR NAME:
LANCE M BARBER
Permit #: 42841
Operator Agency/Dept: HUDSON FALLS VIL. PD
Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:57
INTERNAL STANDARD	VERIFIED	21:57
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	21:58
BLANK TEST	0.00	21:58
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	21:59
BLANK TEST	0.00	22:00
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	22:00
BLANK TEST	0.00	22:01

