

SUPERVISOR TEST

GRANVILLE VILLAGE
POLICE DEPT
DATAMASTER dmt: 113306

Date: 03/22/2022

OPERATOR NAME:
ERNEST R BASSETT
Permit #: 25518
Operator Agency/Dept: GRANVILLE VILLAGE PD
Reference Standard #: 21390

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:05
INTERNAL STANDARD	VERIFIED	15:05
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	15:06
BLANK TEST	0.00	15:07
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	15:07
BLANK TEST	0.00	15:08
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	15:08
BLANK TEST	0.00	15:09

