

SOLUTION CHANGE

CAMBRIDGE VILLAGE
POLICE DEPT
DATAMASTER dmt: 126706

Date: 03/01/2022

OPERATOR NAME:
ROBERT M DANKO
Permit #: 39105
Operator Agency/Dept: CAMBRIDGE VILL. PD
Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:15
INTERNAL STANDARD	VERIFIED	15:15
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	15:15
BLANK TEST	0.00	15:16
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	15:16
BLANK TEST	0.00	15:17
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	15:17
BLANK TEST	0.00	15:18

