

SUPERVISOR TEST

CAMBRIDGE VILLAGE  
POLICE DEPT  
DATAMASTER dmt: 126706

Date: 03/14/2022

OPERATOR NAME:  
ROBERT M DANKO  
Permit #: 39105  
Operator Agency/Dept: CAMBRIDGE VILL. PD  
Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:34
INTERNAL STANDARD	VERIFIED	09:34
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:35
BLANK TEST	0.00	09:36
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:36
BLANK TEST	0.00	09:37
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:37
BLANK TEST	0.00	09:38

