

SUPERVISOR TEST

WASHINGTON COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 106406

Date: 03/07/2022

OPERATOR NAME:
KYLE L KOLAR
Permit #: 44063
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:41
INTERNAL STANDARD	VERIFIED	23:41
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	23:42
BLANK TEST	0.00	23:43
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	23:43
BLANK TEST	0.00	23:44
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	23:44
BLANK TEST	0.00	23:45

B #6a1

