

SUPERVISOR TEST

WASHINGTON COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 106406

Date: 03/14/2022

OPERATOR NAME:
JEFFREY E PETERMAN
Permit #: 49748
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:50
INTERNAL STANDARD	VERIFIED	00:50
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	00:50
BLANK TEST	0.00	00:51
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	00:51
BLANK TEST	0.00	00:52
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	00:52
BLANK TEST	0.00	00:53

