

SUPERVISOR TEST

WASHINGTON COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 106406

Date: 03/21/2022

OPERATOR NAME:
WILLIAM M GREENE
Permit #: 48760
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 21440

Nominal = 0.10 g/210L

| | | |
|-------------------------|----------|-------|
| BLANK TEST | 0.00 | 18:57 |
| INTERNAL STANDARD | VERIFIED | 18:57 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.10 | 18:57 |
| BLANK TEST | 0.00 | 18:58 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.10 | 18:59 |
| BLANK TEST | 0.00 | 18:59 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.10 | 19:00 |
| BLANK TEST | 0.00 | 19:01 |

WMS
#647

