

SOLUTION CHANGE

WASHINGTON CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 100284

Date: 03/11/2022

OPERATOR NAME:  
ROBERT J SULLIVAN  
Permit #: 34106  
Operator Agency/Dept: WASHINGTON CO. SO  
Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:07
INTERNAL STANDARD	VERIFIED	09:08
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:08
BLANK TEST	0.00	09:09
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:09
BLANK TEST	0.00	09:10
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:10
BLANK TEST	0.00	09:11

