

SUPERVISOR TEST

WASHINGTON CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 100284

Date: 03/22/2022

OPERATOR NAME:
ROBERT J SULLIVAN
Permit #: 34106
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:00
INTERNAL STANDARD	VERIFIED	06:00
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:01
BLANK TEST	0.00	06:01
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:02
BLANK TEST	0.00	06:03
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:03
BLANK TEST	0.00	06:04

