

SUPERVISOR TEST

CAMBRIDGE VILLAGE
POLICE DEPT
DATAMASTER dmt: 126706

Date: 02/08/2022

OPERATOR NAME:
ROBERT M DANKO
Permit #: 39105
Operator Agency/Dept: CAMBRIDGE VILL. PD
Reference Standard #: 21390

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:28
INTERNAL STANDARD	VERIFIED	07:28
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	07:28
BLANK TEST	0.00	07:29
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	07:29
BLANK TEST	0.00	07:30
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	07:30
BLANK TEST	0.00	07:31

