

SOLUTION CHANGE

FORT EDWARD
VILLAGE POLICE DEPT.
DATAMASTER dmt: 115708

Date: 03/01/2022

OPERATOR NAME:
JOSEPH P DEANGELO
Permit #: 49218
Operator Agency/Dept: FORT EDWARD VILL. PD
Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:13
INTERNAL STANDARD	VERIFIED	17:13
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	17:14
BLANK TEST	0.00	17:15
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	17:15
BLANK TEST	0.00	17:16
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	17:16
BLANK TEST	0.00	17:17

JPD
#1217

