

SUPERVISOR TEST

FORT EDWARD
VILLAGE POLICE DEPT.
DATAMASTER dmt: 115708

Date: 03/15/2022

OPERATOR NAME:
JOSEPH P DEANGELO
Permit #: 49218
Operator Agency/Dept: FORT EDWARD VILL. PD
Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:18
INTERNAL STANDARD	VERIFIED	23:18
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:18
BLANK TEST	0.00	23:19
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:20
BLANK TEST	0.00	23:21
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:21
BLANK TEST	0.00	23:22

1277

