

SUPERVISOR TEST

-----  
 HUDSON FALLS  
 VILLAGE POLICE DEPT.  
 DATAMASTER dmt: 102408  
 -----

Date: 03/28/2022

OPERATOR NAME:  
 LANCE M BARBER  
 Permit #: 42841  
 Operator Agency/Dept: HUDSON FALLS VIL. PD  
 Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:30
INTERNAL STANDARD	VERIFIED	04:30
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	04:30
BLANK TEST	0.00	04:31
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	04:32
BLANK TEST	0.00	04:33
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	04:33
BLANK TEST	0.00	04:34

