

SUPERVISOR TEST

WHITEHALL VILLAGE  
POLICE DEPT.  
DATAMASTER dmt: 113806

Date: 03/19/2022

OPERATOR NAME:  
ALEXANDER T GILMORE  
Permit #: 47649  
Operator Agency/Dept: WHITEHALL VILL. P.D.  
Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:54
INTERNAL STANDARD	VERIFIED	19:54
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	19:54
BLANK TEST	0.00	19:55
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	19:56
BLANK TEST	0.00	19:56
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	19:57
BLANK TEST	0.00	19:58

