

SUPERVISOR TEST

FORT EDWARD
VILLAGE POLICE DEPT.
DATAMASTER dmt: 115708

Date: 03/29/2022

OPERATOR NAME:
JOSEPH P DEANGELO
Permit #: 49218
Operator Agency/Dept: FORT EDWARD VILL. PD
Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:01
INTERNAL STANDARD	VERIFIED	15:01
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	15:02
BLANK TEST	0.00	15:03
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	15:03
BLANK TEST	0.00	15:04
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	15:04
BLANK TEST	0.00	15:05

JD
#1217

