

SUPERVISOR TEST

-----  
GRANVILLE VILLAGE  
POLICE DEPT  
DATAMASTER dmt: 113306  
-----

Date: 03/13/2022

OPERATOR NAME:  
THOMAS J ZOISTOSKI  
Permit #: 36771  
Operator Agency/Dept: GRANVILLE VILLAGE PD  
Reference Standard #: 21390

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:43
INTERNAL STANDARD	VERIFIED	11:43
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	11:43
BLANK TEST	0.00	11:44
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	11:44
BLANK TEST	0.00	11:45
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	11:45
BLANK TEST	0.00	11:46

