

SOLUTION CHANGE

GRANVILLE VILLAGE
POLICE DEPT
DATAMASTER dmt: 113306

Date: 03/28/2022

OPERATOR NAME:
JASON M VANDENBURGH
Permit #: 37635
Operator Agency/Dept: GRANVILLE VILLAGE PD
Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:42
INTERNAL STANDARD	VERIFIED	13:42
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	13:43
BLANK TEST	0.00	13:43
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	13:44
BLANK TEST	0.00	13:45
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	13:45
BLANK TEST	0.00	13:46

