

SUPERVISOR TEST

GRANVILLE VILLAGE
POLICE DEPT
DATAMASTER dmt: 113306

Date: 04/04/2022

OPERATOR NAME:
JASON M VANDENBURGH
Permit #: 37635
Operator Agency/Dept: GRANVILLE VILLAGE PD
Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:48
INTERNAL STANDARD	VERIFIED	06:48
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	06:48
BLANK TEST	0.00	06:49
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	06:50
BLANK TEST	0.00	06:50
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	06:51
BLANK TEST	0.00	06:51

