

SUPERVISOR TEST

WASHINGTON COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 106406

Date: 03/29/2022

OPERATOR NAME:
KYLE L KOLAR
Permit #: 44063
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:59
INTERNAL STANDARD	VERIFIED	02:59
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	03:00
BLANK TEST	0.00	03:00
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	03:01
BLANK TEST	0.00	03:02
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	03:02
BLANK TEST	0.00	03:03

[Handwritten Signature]
#691

