

SUPERVISOR TEST

-----  
WASHINGTON CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 100284  
-----

Date: 03/28/2022

OPERATOR NAME:  
MATTHEW K JACKSON  
Permit #: 40078  
Operator Agency/Dept: WASHINGTON CO. SO  
Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:35
INTERNAL STANDARD	VERIFIED	04:35
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	04:35
BLANK TEST	0.00	04:36
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	04:36
BLANK TEST	0.00	04:37
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	04:37
BLANK TEST	0.00	04:38

