

SUPERVISOR TEST

HUDSON FALLS
 VILLAGE POLICE DEPT.
 DATAMASTER dmt: 102408

Date: 04/06/2022

OPERATOR NAME:
 LANCE M BARBER
 Permit #: 42841
 Operator Agency/Dept: HUDSON FALLS VIL. PD
 Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:35
INTERNAL STANDARD	VERIFIED	06:35
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	06:35
BLANK TEST	0.00	06:36
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	06:37
BLANK TEST	0.00	06:37
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	06:38
BLANK TEST	0.00	06:39

