

SOLUTION CHANGE

WHITEHALL VILLAGE
POLICE DEPT.
DATAMASTER dmt: 113806

Date: 04/01/2022

OPERATOR NAME:
ALEXANDER T GILMORE
Permit #: 47649
Operator Agency/Dept: WHITEHALL VILL. P.D.
Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:17
INTERNAL STANDARD	VERIFIED	20:17
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	20:17
BLANK TEST	0.00	20:18
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	20:18
BLANK TEST	0.00	20:19
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	20:19
BLANK TEST	0.00	20:20

