

BLANK TEST

WHITEHALL VILLAGE  
POLICE DEPT.  
DATAMASTER dmt: 113806

Date: 04/02/2022

OPERATOR NAME:  
ALEXANDER T GILMORE  
Permit #: 47649  
Operator Agency/Dept: WHITEHALL VILL. P.D.  
Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:49
INTERNAL STANDARD	VERIFIED	20:49
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	20:49
BLANK TEST	0.00	20:50
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	20:50
BLANK TEST	0.00	20:51
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	20:52
BLANK TEST	0.00	20:53

