

WHITEHALL VILLAGE
POLICE DEPT.
DATAMASTER dmt: 113806

Date: 04/09/2022

OPERATOR NAME:
BRYAN J GRECO
Permit #: 44077
Operator Agency/Dept: WHITENALL VILL. P.D.
Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:14
INTERNAL STANDARD	VERIFIED	06:14
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	06:14
BLANK TEST	0.00	06:15
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	06:15
BLANK TEST	0.00	06:16
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	06:17
BLANK TEST	0.00	06:17

