

SUPERVISOR TEST

CAMBRIDGE VILLAGE
POLICE DEPT
DATAMASTER dmt: 126706

Date: 04/18/2022

OPERATOR NAME:
ROBERT M DANKO
Permit #: 39105
Operator Agency/Dept: CAMBRIDGE VILL. PD
Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:28
INTERNAL STANDARD	VERIFIED	08:28
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:29
BLANK TEST	0.00	08:29
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:30
BLANK TEST	0.00	08:31
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:31
BLANK TEST	0.00	08:32

