

SUPERVISOR TEST

GRANVILLE VILLAGE  
POLICE DEPT  
DATAMASTER dmt: 113306

Date: 04/11/2022

OPERATOR NAME:  
JASON M VANDENBURGH  
Permit #: 37635  
Operator Agency/Dept: GRANVILLE VILLAGE PD  
Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:33
INTERNAL STANDARD	VERIFIED	07:33
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	07:33
BLANK TEST	0.00	07:34
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	07:34
BLANK TEST	0.00	07:35
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	07:35
BLANK TEST	0.00	07:36

