

SUPERVISOR TEST

GRANVILLE VILLAGE
POLICE DEPT
DATAMASTER dmt: 113306

Date: 04/18/2022

OPERATOR NAME:
JASON M VANDENBURGH
Permit #: 67635
Operator Agency/Dept: GRANVILLE VILLAGE PD
Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:31
INTERNAL STANDARD	VERIFIED	08:31
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	08:32
BLANK TEST	0.00	08:33
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	08:33
BLANK TEST	0.00	08:34
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	08:34
BLANK TEST	0.00	08:35

