

SUPERVISOR TEST

GRANVILLE VILLAGE
POLICE DEPT
DATAMASTER dmt: 113306

Date: 04/25/2022

OPERATOR NAME:
JASON M VANDENBURGH
Permit #: 67635
Operator Agency/Dept: GRANVILLE VILLAGE PD
Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:54
INTERNAL STANDARD	VERIFIED	10:55
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	10:55
BLANK TEST	0.00	10:56
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	10:56
BLANK TEST	0.00	10:57
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	10:57
BLANK TEST	0.00	10:58

