

SUPERVISOR TEST

WASHINGTON COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 106406

Date: 04/25/2022

OPERATOR NAME:
DALE W QUESNEL
Permit #: 37510
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:12
INTERNAL STANDARD	VERIFIED	14:12
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	14:13
BLANK TEST	0.00	14:14
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	14:14
BLANK TEST	0.00	14:15
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	14:15
BLANK TEST	0.00	14:16

