

SUPERVISOR TEST

WASHINGTON COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 106406

Date: 04/11/2022

OPERATOR NAME:
CORBET W SULLIVAN
Permit #: 45898
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:57
INTERNAL STANDARD	VERIFIED	09:57
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:58
BLANK TEST	0.00	09:59
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:59
BLANK TEST	0.00	10:00
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	10:00
BLANK TEST	0.00	10:02

CS
#623

