

SUPERVISOR TEST

-----  
WASHINGTON CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 100284  
-----

Date: 04/19/2022

OPERATOR NAME:  
TREVOR J PECK  
Permit #: 49747  
Operator Agency/Dept: WASHINGTON CO. SO  
Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:41
INTERNAL STANDARD	VERIFIED	20:41
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	20:42
BLANK TEST	0.00	20:43
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	20:43
BLANK TEST	0.00	20:44
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	20:44
BLANK TEST	0.00	20:45

#637 TJP

