

SUPERVISOR TEST

-----  
FORT EDWARD  
VILLAGE POLICE DEPT.  
DATAMASTER dmt: 115708  
-----

Date: 04/26/2022

OPERATOR NAME:  
JOSEPH P DEANGELO  
Permit #: 49218  
Operator Agency/Dept: FORT EDWARD VILL. PD  
Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:24
INTERNAL STANDARD	VERIFIED	19:24
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	19:25
BLANK TEST	0.00	19:26
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	19:26
BLANK TEST	0.00	19:27
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	19:27
BLANK TEST	0.00	19:28

JD  
#1203

