

SOLUTION CHANGE

FORT EDWARD
VILLAGE POLICE DEPT.
DATAMASTER dmt: 115708

Date: 04/05/2022

OPERATOR NAME:
JOSEPH P DEANGELO
Permit #: 49218
Operator Agency/Dept: FORT EDWARD VILL. PD
Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:39
INTERNAL STANDARD	VERIFIED	15:39
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	15:40
BLANK TEST	0.00	15:41
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	15:41
BLANK TEST	0.00	15:42
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	15:42
BLANK TEST	0.00	15:43

JD
#1217

