

SUPERVISOR TEST

HUDSON FALLS
 VILLAGE POLICE DEPT.
 DATAMASTER dmt: 102408

Date: 04/13/2022

OPERATOR NAME:
 LANCE M BARBER
 Permit #: 42841
 Operator Agency/Dept: HUDSON FALLS VIL. PD
 Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	16:39
INTERNAL STANDARD	VERIFIED	16:39
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	16:39
BLANK TEST	0.00	16:40
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	16:40
BLANK TEST	0.00	16:41
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	16:41
BLANK TEST	0.00	16:42

