

SUPERVISOR TEST

HUDSON FALLS
 VILLAGE POLICE DEPT.
 DATAMASTER dmt: 102408

Date: 04/20/2022

OPERATOR NAME:
 LANCE M BARBER
 Permit #: 42841
 Operator Agency/Dept: HUDSON FALLS VIL. PD
 Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:35
INTERNAL STANDARD	VERIFIED	01:35
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	01:35
BLANK TEST	0.00	01:36
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	01:36
BLANK TEST	0.00	01:37
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	01:38
BLANK TEST	0.00	01:39

