

SOLUTION CHANGE

HUDSON FALLS
 VILLAGE POLICE DEPT.
 DATAMASTER dmt: 102408

Date: 05/01/2022

OPERATOR NAME:
 EVAN J HOAGLAND
 Permit #: 49911
 Operator Agency/Dept: HUDSON FALLS VIL. PD
 Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:58
INTERNAL STANDARD	VERIFIED	22:58
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	22:59
BLANK TEST	0.00	22:59
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	23:00
BLANK TEST	0.00	23:01
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	23:01
BLANK TEST	0.00	23:02

