

SUPERVISOR TEST

HUDSON FALLS
VILLAGE POLICE DEPT.
DATAMASTER dmt: 102408

Date: 05/13/2022

OPERATOR NAME:
LANCE M BARBER
Permit #: 42841
Operator Agency/Dept: HUDSON FALLS VIL. PD
Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:18
INTERNAL STANDARD	VERIFIED	03:18
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	03:19
BLANK TEST	0.00	03:20
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	03:20
BLANK TEST	0.00	03:21
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	03:21
BLANK TEST	0.00	03:22

